



Certificate of Accreditation

This is to certify that

BV Global Limited

56A, Bourdillon Road, Ikoyi, Lagos State, Nigeria

Has been assessed by the Nigeria National Accreditation Service and meets the requirements of the International Standard

ISO/IEC 17025:2005

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation Service:



Issue date: **April 08, 2019**
Validity through: **December 01, 2020**

Scope of Accreditation
Calibration Laboratory
BV Global Limited
56A, Bourdillon Road, Ikoyi, Lagos State, Nigeria

Schedule NO.:C0003

Issue No. 1: 08 04, 2019

Valid to : 01 12, 2020

Measured Quantity	Range	Calbration and Measurement Capability (CMC)	Brief description of calibration method	Brief description of calibration equipment
Gauge Pressure (Pressure gauge Indicator)	0 to 100 bar	±10 Bar	DKD-R6-1:2003	Calibration of a guage pressure against a reference pressure gauge(Hydraulic)