



Certificate of Accreditation

This is to certify that

SELFA (Nigeria) Limited

10, Atunwa Street, Ikeja, Lagos State, Nigeria

Has been assessed by the Nigeria National Accreditation Service and meets the requirements of the International Standard

ISO/IEC 17025:2005

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation Service:



Issue date: **January 21, 2019**
Validity through: **January 20, 2021**

Scope of Accreditation
Calibration Laboratory
SELFA (Nigeria) Limited
10, Atunwa Street, Ikeja, Lagos State, Nigeria

Schedule NO.:C0002

Issue No. 1: 21 01, 2019

Valid to : 20 01, 2021

Measured Quantity	Range	Calibration and Measurement Capability (CMC)	Brief description of calibration method	Brief description of calibration equipment
Gauge pressure (gauge Pressure indicator)	0 to 20 bar	± 0.5bar	ASME B40.100.2005	Calibration of gauge pressure against a reference pressure gauge (hydraulic)