



Type: Initial Accreditation



Certificate of Accreditation

This is to certify that

BIOLOGIX SUPPORT SERVICES LIMITED

56C OLORUNLOGBON STREET, ANTHONY VILLAGE LAGOS, NIGERIA

Has been assessed by the Nigeria National Accreditation System and meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Calibration
No. C0007

Issue date: January 30, 2023
Valid through: January 29, 2025

Scope of Accreditation

BIOLOGIX SUPPORT SERVICES LIMITED

56C OLORUNLOGBON STREET, ANTHONY VILLAGE LAGOS, NIGERIA

Schedule NO.:C0007

Issue No. 1: 30 01, 2023

Valid to : 29 01, 2025

Measured Quantity	Range	Calibration and Measurement Capability (CMC)	Brief description of calibration method	Brief description of calibration equipment
Micropipettes Single and Multi Channels Microliters (ul)	5000ul	± 5.61ul	Gravimetric Method	▪ Analytical Balance ▪ Hygrometer
	100 - 1000ul	± 1.85ul		
	10 - 100ul	±0.05ul		
	50ul	±0.02ul		
	1 - 10ul	±0.01ul		