



Type: Re-Accreditation



Certificate of Accreditation

This is to certify that

BV GLOBAL LIMITED

56A BOURDILLON ROAD IKOYI LAGOS

Has been assessed by the Nigeria National Accreditation System and meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Calibration
No. C0003

Issue date: **May 11, 2023**
Valid through: **May 10, 2025**

Scope of Accreditation
Calibration Laboratory
BV GLOBAL LIMITED
56A BOURDILLON ROAD IKOYI LAGOS

Schedule NO.:C0003A

Issue No. 2: 11 05, 2023

Valid to : 10 05, 2025

Measured Quantity	Range	Calbration and Measurement Capability (CMC)	Brief description of calibration method	Brief description of calibration equipment
PRESSURE	0-1000 BAR	± 1%	DKD	<ul style="list-style-type: none"> ▪ MASTER GAUGE ▪ HAND PUMP ▪ PRESSURE COMPARATOR
TEMPERATURE	-5 TO 500 ⁰	± 0.5%	EURAMET CG-11, EURAMET CG-13	WATER BATH TEMPERATURE BATH PROCESS CALIBRATOR
ELECTRICAL	1000A, 1000V, 100M OhMS	± 0.5%	EURAMET CG-15	PROCESS CALIBRATOR MULTIMETER DECADE BOX OSCILOSCOPE
	999M OhMS	± 0.5%		
	10M Hz	± 1%		
MECHANICAL	0-1000 NM	1%	EURAMET CG-14	TORQUE TESTER KNOWN WEIGHT DIMENSION TOOL
	500 TON	± 0.01%		
		± 1.5%		