



Certificate of Accreditation

This is to certify that

INSTRUMENTATION AND CALIBRATION SERVICES LIMITED

PLOT 8 EASTERN BYPASS WAY, PORT HARCOURT, RIVERS STATE, NIGERIA

Has been assessed by the Nigeria National Accreditation System and
meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Issue date: October 1, 2025
Valid through: September 30, 2027

Scope of Accreditation

Calibration Laboratory

INSTRUMENTATION AND CALIBRATION SERVICES LIMITED

PLOT 8 EASTERN BYPASS WAY, PORT HARCOURT, RIVERS STATE, NIGERIA

Schedule NO.: C0067

Issue No. 1: 1 10, 2025

Valid to : 30 10, 2027

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity	Range	Expanded Measurement Uncertainty (k = 2)	Remark (Calibration Method)
Electrical			
Voltage	0 - 1000V ac/dc	0.002% +/- 1 Digit	Direct Comparison in line with Manufacturers specification
Current	0 - 10A ac/dc	0.02% +/- 2 Digit	Direct Comparison in line with Manufacturers specification
Insulation Resistance	0 - 1GΩ	1% +/- 1 Digit	Direct Comparison in line with Manufacturers specification
Pressure			
Pressure	0 - 31 bar	0.1%	Direct Comparison in line with DKD-R6-1
	0 - 1000 bar	0.05%	Direct Comparison in line with DKD-R6-1
Temperature			
Temperature	-30 to 165 °C	0.1K at -30 °C 0.16K at 165 °C	Direct Comparison in line with Euramet CG-11