



Certificate of Accreditation

This is to certify that

BIOSAFE EQUIPMENT CALIBRATIONS LIMITED

3RD FLOOR OCEAN CENTRE, PLOT 1018 CADASTRAL ZONE B18, OFF OLADIPO
DIYA ROAD, GUDU ABUJA

Has been assessed by the Nigeria National Accreditation System and
meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Calibration
No. C0013

Date of Issue of First Accreditation: **February 16th, 2026**
Date of Issue of Latest Accreditation: **February 16th, 2026**
Expiry Date: **February 15th, 2030**
Reaccreditation Date: **August 15th, 2029**

Scope of Accreditation

Calibration Laboratory

BIOSAFE EQUIPMENT CALIBRATIONS LIMITED

3RD FLOOR OCEAN CENTRE, PLOT 1018 CADASTRAL ZONE B18, OFF OLADIPO DIYA ROAD, GUDU ABUJA

Reference No.: C0008

Issue No. 1: 16 02, 2026

Valid to : 15 02, 2030

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity	Range	Expanded Measurement Uncertainty (k = 2)	Remark (Calibration Method)
1. Mass	1g to 1000g	0.038	(OIML R111-1 2004)
2. Temperature	350 to -86	$\pm 0.03^{\circ}$	DIN EN 60751) BN EN 60068-3-7)
3. Volume	1ul to 1000ul	0.0200	ISO 8566-6)
4. speed	100rpm -90000rpm	0.045	IEC-61010-2-020)