



Certificate of Accreditation

This is to certify that

PETROLEUM TESTING LABORATORY Q & Q CONTROL SERVICES

PLOT 164C IDIAGBON CLOSE, GRA PHASE 3, PORT HARCOURT
Has been assessed by the Nigeria National Accreditation System and
meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Testing

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Testing
No. T0028

Date of Issue of First Accreditation: **January 3rd, 2022**
Date of Issue of Latest Accreditation: **January 30th, 2026**
Expiry Date: **January 29th, 2030**
Reaccreditation Date: **July 29th, 2029**

Scope of Accreditation

Chemical Testing Laboratory

PETROLEUM TESTING LABORATORY Q & Q CONTROL SERVICES

PLOT 164C IDIAGBON CLOSE, GRA PHASE 3, PORT HARCOURT

Reference No.: TL0010D

Issue No. 2: 30 10, 2026

Valid to : 29 01, 2030

| Material or Products Tested | Types of Tests/ Properties Measured / Range of Measurements | Standard Specifications, Techniques/Equipment Used |
|---|---|--|
| Gasoil, Fuel Oil | Ash Content | ASTM D482/ Ash Content Unit |
| Gasoline, Gasoil, DPK, JET A-1, Fuel Oil, Crude Oil | Density @15°C/ Relative Density @60/60°F and API | ASTM D1298/Hydrometer |
| Fuel Oil, Gasoil, Gasoline | Density @15°C/ Relative Density @60/60°F/API | ASTM D4052/Densitometer |
| Gasoline, Gasoil, DPK, JET A-1 | Distillation (Atmospheric) | ASTM D86/ Distillation Unit |
| Gasoil, Fuel Oil | Flash Point (PMCC) | ASTM D93/PMCC |
| Gasoil, Fuel Oil, Crude oil | Sediment by Extraction | ASTM D473/Sediment by Extraction Unit |
| Gasoil, Fuel Oil, Crude Oil | Viscosity, Kinematic @ 37.8°C, 40°C, 50°C | ASTM D445/Viscosity Unit |
| Gasoil, Fuel Oil | Water by Distillation | ASTM D95 |