



Certificate of Accreditation

This is to certify that

BV GLOBAL LIMITED

PLOT 9 JC STREET, OFF PETER ODILI ROAD, TRANS AMADI
LAYOUT, PORT HARCOURT, RIVERS STATE

Has been assessed by the Nigeria National Accreditation System and
meets the requirements of the International Standard

ISO/IEC 17025:2017

with demonstrated technical competence in the field of

Calibration

for the specific services listed on the approved Scope of Accreditation.

On behalf of the Nigeria National Accreditation System:



Calibration
No. C0003

Date of Issue of First Accreditation: **March 2rd, 2026**
Date of Issue of Latest Accreditation: **March 2rd, 2026**
Expiry Date: **March 1st, 2030**
Reaccreditation Date: **September 1st, 2029**

Scope of Accreditation

Calibration Laboratory

BV GLOBAL LIMITED

PLOT 9 JC STREET, OFF PETER ODILI ROAD, TRANS AMADI LAYOUT, PORT HARCOURT, RIVERS STATE

Reference No.: CL00006

Issue No. 1: 02 03, 2026

Valid to : 01 03, 2030

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity	Range	Expanded Measurement Uncertainty (k = 2)	Remark (Calibration Method)
1. Pressure	0-1000 bar	±1%	DKD
2. Temperature	-5 -500 °C	±0.5%	EURAMETCG-11 EURAMET CG-13
3. Electrical			
DC voltage	0 mV - 900 mV	±0.5%	EURAMETCG-15
DC current	0 mA - 18 mA,	±0.5%	EURAMETCG-15
Resistance	0 Ohm - 9.9 kOhms	±0.5%	EURAMETCG-15
Frequency	0 Hz - 10 MHz	±1%	EURAMETCG-15
4. Mechanical (Torque)	0-1000 Nm	1%	EURAMETCG-14